

JEOL 1200

JEOL 1200 Detailed Status

Blog stream

Create a blog post to share news and announcements with your team and company.



UP

This scope is fully operational enjoy.

Loose connector on the IS shifts and PLA shifts has been repaired.



Checked diffusion pumps. There are no shorts. Checked vacuum PB and discovered a bad connection. System now up and running.... enjoy!